



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/690,436	10/22/2003	Sharon Mi Lyn Tan	12712/46001	3696
23838	7590	09/09/2005	EXAMINER	
KENYON & KENYON 1500 K STREET NW SUITE 700 WASHINGTON, DC 20005			THANH, LOAN H	
			ART UNIT	PAPER NUMBER
			3763	

DATE MAILED: 09/09/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/690,436

Applicant(s)

TAN, SHARON MI LYN

Examiner

LoAn H. Thanh

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 22 October 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 22 October 2003 is/are: a) ☐ accepted or b) ☒ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date: _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date <u>03/21/05</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Drawings

Figure 1 should be designated by a legend such as --Prior Art-- because only that which is old is illustrated. See MPEP § 608.02(g). Corrected drawings in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. The replacement sheet(s) should be labeled "Replacement Sheet" in the page header (as per 37 CFR 1.84(c)) so as not to obstruct any portion of the drawing figures. If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the region of catheter comprising a material which permits passage of an antimicrobial agent release from the intervention device from the catheter lumen to an outer surface of the catheter (cls. 6 & 15), the cap is a stopper (cls. 9 & 18) must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

The drawings are objected to because the threaded portion of the cap and sidearm is not shown or labeled with a reference numeral (cls. 8 & 17).

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended

Art Unit: 3763

replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Liu et al. (USPN 6,371,944) in view of Davidson (USPN 5,588,443).

Liu et al. disclose a system for use in the body comprising a catheter 7, a side arm, a lumen through the side arm 9, which communicates with the catheter lumen and

Art Unit: 3763

a one-way valve 5 and an intervention device such as the wire 18. Liu et al. however does not teach the wire to be coated with an antimicrobial agent. Davidson discloses a wire for use in the body, which is coated with antimicrobial, anticoagulants or, medicaments as desirable or for reducing adhesion or other adverse cellular or tissue response to surfaces in contact with blood. It would have been obvious to one of ordinary skill in the catheter art at the time the invention was made to modify the wire used to contact with blood of Liu et al. with an antimicrobial coating as taught by Davidson in order to prevent infections within the internal system of the patient.

With respect to claims 5 and 14, it is well known in the surgical and medical arts to prepare/swab a patient with iodine before injecting or surgically introducing a device into a patient. With that knowledge in mind, it would have been obvious to one of ordinary skill to swab/coat a device with Iodine as a mere obvious selection of medicament which would prevent infection.

With respect to claims 7-9 and 16-18, Liu et al does not disclose a threaded or stopper cap with the rod affixed to the cap. Fischell et al. disclose a sealing cap 80 for sealing the system to reduce the blood loss when inserting guiding catheters/rods/wires. See figures 1-3 which discloses a stopper cap and figure 7 which disclose a threaded cap. It would have been obvious to one of ordinary skill in the medical art to modify the device of Liu et al. with an end cap for the side arm as taught by Fischell et al. in order to reduce blood loss when inserting the wire into the side arm.


Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to LoAn H. Thanh whose telephone number is (571) 272-4966. The examiner can normally be reached on Mon. - Fri. (First Friday off).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick Lucchesi can be reached on (571) 272-4977. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).


LoAn H. Thanh
Primary Examiner
Art Unit 3763

8/31/05

LT